




CERTIFICATE OF ANALYSIS			
DEPARTMENT	:	QUALITY CONTROL DEPARTMENT	
PRODUCT NAME	:	Cyclone Disposable Three Layer Surgical Mask	
Batch. No.	:	FM/1220	A.R.NO. : 20/A/137
Batch Size	:	200000 Nos.	Pack Size : 1 X 50 pcs
Mfg. Date	:	04/2020	Exp. Date : 03/2024
Date of receipt	:	12/04/2020	Date of Analysis : 16/04/2020

Sr. No.	TESTS	SPECIFICATIONS	RESULTS
1.	Description	Rectangular face mask with a shapeable nose-strip and four ear loop present, two on each side, in order to hold mask in place.	Rectangular face mask with a shapeable nose-strip and four ear loop present, two on each side, in order to hold mask in place.
2.	Material Description	1. Outer Material: Spunbond Polypropylene (Non-Woven Fabric) (SBPP) 20GSM / 25GSM 2. Filter Layer : 95 to 99% Bacterial Efficiency Filter 3. Inner material : Spunbond Polypropylene(Non-Woven Fabric) (SBPP) 20 GSM	Complies with given specification
3.	Ply Nature	3 Ply construction (Non Oven Cotton Layer with Bacterial Filter, 1 paper 1 filter 1 non oven 3 layer)	3 Ply construction (Non Oven Cotton Layer with Bacterial Filter, 1 paper 1 filter 1 non oven 3 layer)
4.	Nose Strip	Plastic Coated / Powder Coated Iron	Complies with specification
5.	Length of Ear Loop	120mm to 140mm	Complies with specification
6.	Length of Nose-strip	63.5mm to 85mm	Complies with specification
7.	Size	Length 175mm & Width 85mm to 95mm	Complies with specification
8.	Bacterial Filtration Efficiency (BFE) count	BFE @3.0µm large Bacteria \geq 99%	BFE @3.0µm large Bacteria \geq 97.5%
9.	Pack size	Pack of 50pcs	Pack of 100 pcs

Remarks: The sample referred complies comply with prescribed standard quality with respect to above test. The product is of standard quality.

CE Mark Certified Product: CE Certificate No. 8577

	PREPARED BY	CHECKED BY	APPROVED BY
SIGN			
DATE	17/02/2020	18/02/2020	19/02/2020
DESIGNATION	QC Executive	Head QC	Head QA